


Technical Documentation	Neutral Electrodes	TD05_02-05-001
	DECLARATION OF COMPATIBILITY SWAROPLATE	ISSUE: 003
		PAGE: 1 OF 3

According to the state of knowledge at the revision date of this Declaration of Compatibility the company Nissha Medical Technologies Ltd. declares that the products of the product family neutral electrodes are compatible to high frequency surgery generators (HF generators) referred to the following list:


AESCULAP
<u>HF generators:</u> GN 300, GN 640 (Nelson deluxe)
<u>Non-split electrodes REF:</u> 1202H, 1203H, 1210H ⁽³⁾ , 1268H ⁽³⁾
<u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1223H, 1226H, 1280H, 1283H, 1269H, 1274H, 1286H

BAYLIS
<u>HF generators:</u> RFP-100A
<u>Non-split electrodes REF:</u> -
<u>Split electrodes REF:</u> 1212H, 1255H

BOWA
<u>HF generators:</u> ARC 400, ARC 350, ARC 303, ARC 250, ARC 100
<u>Non-split electrodes REF:</u> 1202H, 1203H, 1210H ⁽³⁾ , 1258H ⁽³⁾ , 1262H, 1263H
<u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1223H, 1226H, 1280H, 1283H, 1251H, 1252H, 1253H, 1255H, 1257H, 1259H, 1285H, 1288H

CONMED
<u>HF generators:</u> System 2450, System 5000, BiCap III
<u>Non-split electrodes REF:</u> 1202H, 1203H, 1210H ⁽³⁾ , 1258H ⁽³⁾ , 1262H, 1263H
<u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1223H, 1226H, 1280H, 1283H, 1251H, 1252H, 1253H, 1255H, 1257H, 1259H, 1285H, 1288H

ERBE
<u>HF generators:</u> VIO 3, VIO 300 D, VIO 300 S, VIO 200 D, VIO 200 S, ICC 350
<u>Non-split electrodes REF:</u> 1202H, 1203H, 1210H ⁽³⁾ , 1258H ⁽¹⁺³⁾ , 1262H ⁽¹⁾ , 1263H ⁽¹⁾
<u>Split electrodes REF:</u> 1212H, 1213H, 1220H, 1226H, 1280H ⁽²⁾ , 1283H, 1255H ⁽¹⁾ , 1257H ⁽¹⁾ , 1259H ⁽¹⁾ , 1285H ⁽¹⁺²⁾ , 1288H ⁽¹⁾

Technical Documentation	Neutral Electrodes	TD05_02-05-001
	DECLARATION OF COMPATIBILITY SWAROPLATE	ISSUE: 003
		PAGE: 2 OF 3


HEBU
<u>HF generators:</u> HBS x-touch 300 <u>Non-split electrodes REF:</u> 1202H, 1203H, 1210H ⁽³⁾ , 1268H ⁽³⁾ <u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1223H, 1226H, 1280H, 1283H, 1269H, 1274H, 1286H

MEYER-HAAKE
<u>HF generators:</u> radioSURG 2200m, 2200 „TP“, 2200 „TPA“ <u>Non-split electrodes REF:</u> 1201H, 1202H, 1203H, 1210H, 1268H, 1271H <u>Split electrodes REF:</u> -

OLYMPUS
<u>HF generators:</u> ESG-400 <u>Non-split electrodes REF:</u> 1202H, 1203H, 1210H ⁽³⁾ , 1258H ⁽³⁾ , 1262H, 1263H <u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1223H, 1226H, 1280H, 1283H, 1251H, 1252H, 1253H, 1255H, 1257H, 1259H, 1285H, 1288H

RICHARD WOLF
<u>HF generators:</u> 2260 <u>Non-split electrodes REF:</u> - <u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1226H, 1280H, 1251H, 1252H, 1255H, 1257H, 1259H, 1285H,

STORZ
<u>HF generators:</u> AUTOCON III 400 <u>Non-split electrodes REF:</u> 1202H, 1203H, 1210H ⁽³⁾ , 1258H ⁽³⁾ , 1262H, 1263H <u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1223H, 1226H, 1280H, 1283H, 1251H, 1252H, 1253H, 1255H, 1257H, 1259H, 1285H, 1288H

Technical Documentation	Neutral Electrodes	TD05_02-05-001
	DECLARATION OF COMPATIBILITY SWAROPLATE	ISSUE: 003
		PAGE: 3 OF 3

SUTTER
<u>HF generators:</u> CURIS <u>Non-split electrodes REF:</u> 1202H, 1203H, 1210H ⁽³⁾ , 1268H ⁽³⁾ <u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1223H, 1226H, 1280H, 1283H, 1269H, 1274H, 1286H

VALLEYLAB
<u>HF-generators:</u> Force FX, Force EZ, Force Triad, FT10 Energy Platform <u>Non-split electrodes REF:</u> - <u>Split electrodes REF:</u> 1211H, 1212H, 1213H, 1220H, 1222H, 1223H, 1226H, 1280H, 1283H, 1251H, 1252H, 1253H, 1255H, 1257H, 1259H, 1285H, 1288H

Notes:

- (1) International devices only
- (2) NOT compatible with ICC 350
- (3) Allowed to be used with Children only

The compatibility with the listed HF-generators is also given if the original connection cables from the manufacturer of the HF-generators are used and if they support a connection lug of 24 mm in width.

Warning:

If the HF-generator has a contact quality monitoring system, (CQM, e.g. “REM,” “NESSY,” “EASY”) always use split electrodes. Only when a split electrode is used with a CQM system, the contact with the patient is monitored automatically.

Nissha Medical Technologies Ltd.



Daren Davies
Regulatory Compliance Manager

Paignton, 2023-11-01